

 <b>University Health™</b>	Policy #: Rad Proc 13.14.7
SUBJECT: CT Abdomen/Pelvis Oncology 6.7	Effective: 10/2013 Reviewed: 03/07/2017
APPROVED BY: <b>Body Imaging Medical Director</b>	Page 1 of 3

**Purpose:** To provide computed tomography staff with the required protocol for performing CT Abdomen/Pelvis for Oncology Protocol

**Scope:** All adult patients 18 years and older.

**Clinical Indication:** Neoplasm

**Patient Preparation:** Clear liquid diet (6 hours prior to exam)

**Orientation:** Feet first

**Breathing:** Inspiration

**Oral Contrast:** 800-1000ml 90-120 minute drinking period (see proc. 6.12.1)

**IV Contrast per Weight:** 1ml/lb or 2ml/kg not to exceed 150ml injected @ 3ml/sec

**Coverage:** Dome of diaphragm through iliac crest; for pelvis continue through ischial tuberosities

**Anatomic Reference:** Xiphoid process;

**Scan Delay:** Smart prep or care bolus

**Group 1: Non-contrast – Dome of diaphragm to iliac crest**

Scan Mode	Thickness Speed Pitch	Table Interval	SFOV	kVp	Auto mA/ Noise Index	Prep Time (sec)	Recon Type
LS 16 0.8 sec Helical Full	5 27.50 1.375:1	5	Large	120	80-440 11.5	N	Standard
VCT 0.8 sec Helical Full	5 55.00 1.375:1	5	Large	120	80-600 11.5	N	Standard
AS 64 0.5 sec Helical	5 1	5	360	120	Ref MAS 250	N	Standard

 <b>University Health™</b>	<b>Policy #: Rad Proc 13.14.7</b>
<b>SUBJECT: CT Abdomen/Pelvis Oncology 6.7</b>	<b>Effective: 10/2013 Reviewed: 03/07/2017</b>
<b>APPROVED BY: Body Imaging Medical Director</b>	<b>Page 2 of 3</b>

Group 2: Venous Abdomen and pelvis– Dome of diaphragm through ischial tuberosities.

**Smart prep(GE) or care bolus(Siemens) liver: Monitoring delay 45 sec; ISD 5 sec: threshold 50**

Scan Mode	Thickness Speed Pitch	Table Interval	SFOV	kVp	Auto mA/ Noise Index	Prep Time (sec)	Recon Type
<b>LS 16 0.8 sec Helical Full</b>	5 27.50 1.375:1	5	Large	120	80-440 11.5	Smart Prep Liver	Standard
<b>VCT 0.8 sec Helical Full</b>	5 55.00 1.375:1	5	Large	120	80-600 11.5	Smart Prep Liver	Standard
<b>AS 64 0.5 sec Helical</b>	5 1	5	360	120	Ref MAS 250	Care Bolus Liver	Standard

**Group 3 Five minute delays- through kidneys only**

Scan Mode	Thickness Speed Pitch	Table Interval	SFOV	kVp	Auto mA/ Noise Index	Prep Time (sec)	Recon Type
LS 16 0.8 sec Helical Full	5 27.50 1.375:1	5	Large	120	80-440 11.5	5 min	Standard
VCT 0.8 sec Helical Full	5 55.00 1.375:1	5	Large	120	80-600 11.5	5 min	Standard

 <b>University Health™</b>	<b>Policy #: Rad Proc 13.14.7</b>
<b>SUBJECT: CT Abdomen/Pelvis Oncology 6.7</b>	<b>Effective: 10/2013 Reviewed: 03/07/2017</b>
<b>APPROVED BY: Body Imaging Medical Director</b>	<b>Page 3 of 3</b>

AS 64 0.5 sec Helical	5 1	5	360	120	Ref MAS 250	5 min	Standard
-----------------------------	--------	---	-----	-----	----------------	-------	----------

**Algorithm: Recon 1 & 2 Standard**

**Reformation: 2.0 Sagittal and Coronal reformations of the venous phase.**

**Network: Recon 1 (5mm) to PACS. Recon 2 (1.25) auto transmit to AWSERV**

**Notes: Auto mA must be on and mA table checked prior to scanning.**