

	Policy #: Rad Policy 14.4.5
SUBJECT: MR Safety - Prisoners or Patients with Metallic Devices	Effective: 10/1/2018 Revised: 2/2019
APPROVED BY Eduardo Gonzalez-Toledo, MD PhD	Page 1 of 1

Purpose: To define the criteria and processes to be used for screening a patient prior to entering the MR scan room.

Policy:

1. Patients, prisoners or parolees with metallic prisoner-restraining devices or RF ID/tracking devices must have such devices removed prior to entering the Zone 4 of the MR environment.
2. The appropriate authorities accompanying the patient will remove metallic device and replace with disposable single strap cuffs prior to patient entering into Zone 4.
3. The appropriate authorities accompanying the patient will replace original device immediately upon completion of the MR procedure.

