

	POLICY #: Rad Policy 14.8.1
SUBJECT: : Cardiac Pacemaker and Implanted Cardiac Stimulators	Effective: 10/1/2018 Revised 02/2019
APPROVED BY Eduardo Gonzalez-Toledo, MD PhD	Page 1 of 2

PURPOSE: To describe the process that should be followed should a patient enter Zone 3 or 4 with an ICD or pacemaker.

POLICY:

1. Pt should be slowly removed from the magnet.
2. MR Medical Director and MR Manager should be notified.
3. Patient's physician should be notified.
4. Cardiology should be notified.
5. Pt will be taken to the heart station so that a representative can interrogate the device for adverse effects.
6. A variance will be submitted.

	POLICY #: Rad Policy 14.8.1
SUBJECT: : Cardiac Pacemaker and Implanted Cardiac Stimulators	Effective: 10/1/2018 Revised 02/2019
APPROVED BY Eduardo Gonzalez-Toledo, MD Phd	Page 1 of 2